

A novel phenotypic detection strategy for class A, B and OXA-48 carbapenemases in Enterobacteriaceae using temocillin

Karin van Dijk^{1,2}, Jelle Scharringa², Guido Voets², Sebastiaan Voskuil², Ad Fluit², Wouter Rottier², Maurine Leverstein-Van Hall^{3,4}, James Cohen Stuart^{5,6}

Departments of Medical Microbiology of ¹Academic Medical Center, Amsterdam, ²University Medical Center Utrecht, Utrecht and ³Bronovo Hospital, The Hague
⁴Diaconessenhuis, Leiden ⁵Regional Laboratory of Public Health, Haarlem ⁶Medical Center Alkmaar, Alkmaar

Introduction

Class A and B carbapenemases in Enterobacteriaceae may be detected using carbapenemase inhibition tests with boronic acid derivatives (BA) and dipicolinic acid (DPA)/EDTA, respectively. However, for OXA-48 (like) carbapenemases, no specific inhibitor is available. Since OXA-48 confers high-level temocillin resistance, a disc diffusion assay using temocillin besides BA and DPA inhibition tests was evaluated for detection of class A, B and OXA-48 carbapenemases.

Methods I


The test collection included 128 well characterized non-repeat Enterobacteriaceae isolates suspected for carbapenemase production, i.e. with meropenem MIC \geq 0.5 mg/L (83 *K. pneumoniae*, 17 *E. coli*, 23 *Enterobacter* spp., 3 *P. mirabilis*, 2 *S. marcescens*). The isolates produced KPC (n=36), MBL (n=31), KPC plus MBL (n=4), OXA-48 (n=25), OXA-162 (n=2), ESBL (n=19), AmpC (n=10) or ESBL plus AmpC (n=1). PCR and sequencing of beta-lactamase genes was used as reference test. Phenotypic carbapenemase detection was performed with discs (Rosco) containing meropenem (10 ug), temocillin (30 ug), meropenem + phenyl BA (PBA), meropenem + DPA, meropenem + PBA + DPA, and meropenem + cloxacillin (CL).

Methods II

Table 1 shows the strategy for interpretation of carbapenemase inhibition tests. First, to identify KPC and MBL producers, inhibition tests with PBA and/or DPA and CL were evaluated. Second, when no synergy between meropenem and PBA, DPA or both was observed, absence of an inhibition zone (\leq 10 mm) around the temocillin disc was used to identify OXA carbapenemases.

Results

For identification of class A, B and OXA carbapenemases the sensitivity was 97%, 90% and 100%, respectively. Due to swarming, interpretation of two *P. mirabilis* isolates was false negative. Sensitivity for class B detection in non *Proteus* spp. was 97%. The sensitivity for all classes was 96% (98% in non *Proteus* spp.). None of the 27 OXA producers showed an inhibition zone around the temocillin disc. ESBL and/or AmpC producers had temocillin zone diameters between 13 to 29 mm.



Conclusion

A 30 ug temocillin disc with a zone breakpoint of \leq 10 mm added to carbapenemase inhibition tests with PBA and DPA, enables sensitive and specific detection and identification of KPC, MBL and OXA-48, the most prevalent carbapenemases in Enterobacteriaceae.

Phenotypic confirmation tests	Class of Carbapenemase			AmpC with reduced permeability	ESBL with reduced permeability
	Class A	Class B	Class D		
Carbapenem +/- BA	+	-	-	+	-
Carbapenem +/- cloxacillin	-	-	-	+	-
Carbapenem +/- DPA	-	+	-	-	-
Temocillin (zone diameter \leq 10 mm)	-	+/-	+	-	-

BA= boronic acid; DPA= dipicolinic acid

